

Calibration Data Card

SHEAR ACCELEROMETER

MODEL #: 353B33

SERIAL #: 138075

SENSITIVITY: 101.1 mV/g
(10.31 mV/m/s²)

BIAS LEVEL: 10.4 V

Date: 1/14/10 By: Joseph Rogerson

*For further information, please refer to the
accompanying calibration certificate.*

PCB PIEZOTRONICS™
3425 WALDEN AVE. DEPEW, NY 14043
888-684-0013