



Al Conrad <aconrad@lbto.org>

Fwd: Re: LINC NIRVANA (Stirling)

Ralph Hofferbert <hoffer@mpia.de>

Wed, Feb 11, 2015 at 6:48 AM

To: John Little <jlittle@lbto.org>, LINC NIRVANA <LN@lbto.org>

Dear John,

here is Werner's as-built input regarding the Stirling cooler box.

There is a drawing in the attachment. Please consider the following notes:

- The lid can be removed, such that the overall height is not 1985mm but rather 1900mm, to make it compliant with doors and the elevator.
- The side walls are also removable to avoid collisions with the handles. This results in a width of 960mm.
- The backside wall (the one with the cables and tubes, see the attached photo) is not removable. The holding plate adds about 100mm to the envelope + the bent cables and tubes.
- Werner has organized the following transport rollers <http://www.torwegge.de/components/jlb3-k-vu-plus.html>

Would you need more?

Cheers,
Ralph...

2 attachments**P1010194.JPG**
4085K **11799-00 A Max-Planck-Inst. Genehmigungszeichnung_MV.pdf**
538K