



**ASSURANCE QUALITE**  
QUALITY ASSURANCE

- FICHE D'INCIDENT/ANOMALIE**  
INCIDENT/ANOMALY REPORT
- DEMANDE DE DEROGATION**  
WAIVER REQUEST
- DEMANDE DE DEVIATION**  
DEVIATION REQUEST

N° 4836 Rév.  
Date: 03-05-2011  
Page: 1/ 1  
Réf. Projet/Project #:

Diffusion/Distribution : PC CUSTOMER	Affaire/File : MPI	N° dossier /File # : 4988
Resp. Aff./Project manager : LG	Ensemble/Assembly :	N° cde/PO # :
Original : AQ/QA	Pièce/Part : Elliptical cylindrical LENS	Qté/Qty : 1 N°/# : 2
	Plan/Drawing : 4988 / 100 F001 Ind. : B	Doc. annexes, PV : Associated doc, IR : 89271

**TITRE/TITLE :** FLATNESS in TRANSMISSION (SLOPES)

DOCUMENTS APPLIC. APPL. DOCUMENTS	drawing	FL n°
<input type="checkbox"/> MINEURE/MINOR	<input checked="" type="checkbox"/> MAJEURE/MAJOR	<input type="checkbox"/> CRITIQUE/CRITICAL

Constat / Description :

FLATNESS : (SLOPES)      Spec : Sum of 26 fitted zernike polynomials  
(tilt + cylinder removed.)  
< 9 nm / mm  
measure : 10,4 nm / mm

<b>Traitement proposé / Proposed processing</b>		<b>Actions produit/Action on product</b>		<b>Resp. resp.</b>	<b>Pour le due date</b>	<b>Clos le closed on</b>
<input checked="" type="checkbox"/> Accept° en l'état / Accepted as is Retouche / Correction Réparation / Repair Rebut / Rejection Action sur les causes / action on causes Autre/Other :						
	Rédigé par/Written by	Resp. Aff./Project manager	Resp. Qualité/Quality manager			
NOM / NAME	Combes	L. GUESTIN	M. SEBILLEAU			
Date / Visa	03-05-11	03.05.11	04.05.11			
Commentaire client/Customer comment	Accep./Accep.	Refusé/rejected	Société/Company	Date - Visa		
<b>Clôture traitement anomalie</b>		<b>Persistence anomalie</b>		<b>Resp. Qualité/Quality manager</b>		<b>Observ. /Comment</b>
<b>Non Conformance Close Out</b>		<b>Anomaly persisting</b>		M. SEBILLEAU		
Doc de clôture/Close out doc. :		Oui/Yes	Non/no			

AQSESO/018.FRM1.F